




# CERTIFICATE OF ANALYSIS

## Chromatogram



## Quantitative Results

Detector A

Compound Name	Concentration, %
CBDV	--
CBDA	3.810
CBGA	--
CBG	--
CBD	12.470
THCV	--
CBN	--
THC	0.145
CBC	0.175
THCA-A	--
CBL	--
CBDVA	--
CBDB	--

## Sample information

**Batch number:** Batch 401  
**Date of analysis:** 2022 03 22

## Summary

<b>Total THC</b>	<b>0.14</b>	<b>%</b>
<b>Total THC</b>	<b>1.45</b>	<b>mg/g</b>
<b>Total CBD</b>	<b>15.81</b>	<b>%</b>
<b>Total CBD</b>	<b>158.11</b>	<b>mg/g</b>

-- — compound below detection limit or not detected; LOQ <0.01%  
THC content does not exceed legal limits.

Instrumental and analytical conditions.

Equipment: Quantitative analysis was performed using Shimadzu Cannabis Analyzer for Potency - an integrated HPLC system with built-in sample cooler, degasser, autoinjector and UV detector. NexLeaf CBX for potency, 2.7 µm, 4.6 x 150 mm column coupled with NexLeaf CBXGuard column. Shimadzu LabSolutions software.