




# CERTIFICATE OF ANALYSIS

## Chromatogram



### Quantitative Results

Detector A

Compound Name	Concentration, %
CBDV	0.275
CBDA	2.010
CBGA	1.207
CBG	19.060
CBD	3.277
THCV	--
CBN	--
THC	0.033
CBC	0.246
THCA-A	--
CBL	--
CBDVA	--
CBDB	--

### Sample information

**Sample name:** 2000mg CBG RAW 500 mg CBD RAW  
**Batch number:** Batch 643  
**Date of analysis:** 2022 11 14

### Summary

<b>Total THC</b>	<b>0.03</b>	<b>%</b>
<b>Total THC</b>	<b>0.33</b>	<b>mg/g</b>
<b>Total CBG</b>	<b>20.12</b>	<b>%</b>
<b>Total CBG</b>	<b>201.18</b>	<b>mg/g</b>

-- — compound below detection limit or not detected; LOQ <0.01%  
 THC content does not exceed legal limits.

Instrumental and analytical conditions.

Equipment: Quantitative analysis was performed using Shimadzu Cannabis Analyzer for Potency - an integrated HPLC system with built-in sample cooler, degasser, autoinjector and UV detector. NexLeaf CBX for potency, 2.7 µm, 4.6 x 150 mm column coupled with NexLeaf CBXGuard column. Shimadzu LabSolutions software.